

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080323\
 Data File : VN078749.D
 Acq On : 03 Aug 2023 13:43
 Operator : JC\MD
 Sample : 03884-01DL 10X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 001-WILLETS-PT-BLVD(JUNE)DL

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 08/04/2023
 Supervised By :Semsettin Yesilyurt 08/04/2023

Quant Time: Aug 04 06:30:54 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N080123W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Aug 02 04:51:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.824	128	107911	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	9.112	114	770015	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.870	117	391771	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.588	65	306487	37.890	ug/l	0.00
Spiked Amount 30.000	Range 91 - 110		Recovery = 126.300%#			
60) 4-Bromofluorobenzene	12.859	95	113846	19.390	ug/l	0.00
Spiked Amount 30.000	Range 63 - 112		Recovery = 64.633%			
63) Toluene-d8	10.576	98	582770	30.112	ug/l	0.00
Spiked Amount 30.000	Range 91 - 112		Recovery = 100.367%			
Target Compounds						
15) Acetone	4.441	58	46798	56.566	ug/l	84
25) Chloroform	7.977	83	20658	1.594	ug/l	94
30) 2-Butanone	7.500	43	12714m	2.839	ug/l	
62) Toluene	10.635	91	850171	36.077	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080323\
Data File : VN078749.D
Acq On : 03 Aug 2023 13:43
Operator : JC\MD
Sample : 03884-01DL 10X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
001-WILLETS-PT-BLVD(JUNE)DL

Quant Time: Aug 04 06:30:54 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N080123W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Wed Aug 02 04:51:47 2023
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 08/04/2023
Supervised By :Semsettin Yesilyurt 08/04/2023

