

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080323\
 Data File : VN078750.D
 Acq On : 03 Aug 2023 14:07
 Operator : JC\MD
 Sample : 03884-02DL 10X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 002-35TH-AVE(JUNE)DL

Quant Time: Aug 04 06:31:03 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N080123W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Aug 02 04:51:47 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.818	128	93572	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	9.106	114	580962	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.871	117	402998	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.582	65	226246	32.256	ug/l	0.00
Spiked Amount 30.000	Range	91 - 110	Recovery	= 107.533%		
60) 4-Bromofluorobenzene	12.853	95	173827	28.782	ug/l	0.00
Spiked Amount 30.000	Range	63 - 112	Recovery	= 95.933%		
63) Toluene-d8	10.576	98	581313	29.200	ug/l	0.00
Spiked Amount 30.000	Range	91 - 112	Recovery	= 97.333%		
Target Compounds						Qvalue
15) Acetone	4.442	58	45992	64.110	ug/l	97
25) Chloroform	7.971	83	17077	1.520	ug/l	93
30) 2-Butanone	7.506	43	13606	4.026	ug/l #	74
62) Toluene	10.635	91	831801	34.314	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080323\
Data File : VN078750.D
Acq On : 03 Aug 2023 14:07
Operator : JC\MD
Sample : 03884-02DL 10X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
002-35TH-AVE(JUNE)DL

Quant Time: Aug 04 06:31:03 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N080123W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Wed Aug 02 04:51:47 2023
Response via : Initial Calibration

