

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN080418\
 Data File : VN050362.D
 Acq On : 4 Aug 2018 13:17
 Operator : MD\SY
 Sample : J4328-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-01

Quant Time: Aug 10 08:44:22 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 26 18:10:10 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	635874	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1020084	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	884864	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	387132	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	443344	49.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.94%	
35) Dibromofluoromethane	7.59	113	390825	46.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.84%	
50) Toluene-d8	10.09	98	1416547	46.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.52%	
62) 4-Bromofluorobenzene	12.40	95	428701	42.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.04%	
Target Compounds						
16) Acetone	3.82	43	39385	14.76	ug/l	95
19) Methyl tert-butyl Ether	5.05	73	54202	2.78	ug/l	98
65) Chlorobenzene	11.44	112	137773	6.16	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN080418\
Data File : VN050362.D
Acq On : 4 Aug 2018 13:17
Operator : MD\SY
Sample : J4328-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-01

Quant Time: Aug 10 08:44:22 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072418W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 26 18:10:10 2018
Response via : Initial Calibration

