

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN080418\
 Data File : VN050366.D
 Acq On : 4 Aug 2018 14:57
 Operator : MD\SY
 Sample : J4328-06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-03

Quant Time: Aug 10 09:02:47 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 26 18:10:10 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	632714	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1000562	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	878039	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	395558	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	432941	49.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.08%	
35) Dibromofluoromethane	7.59	113	385586	47.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.40%	
50) Toluene-d8	10.09	98	1404951	46.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.56%	
62) 4-Bromofluorobenzene	12.40	95	430515	43.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.06%	
Target Compounds						
16) Acetone	3.82	43	35474	12.94	ug/l	99
65) Chlorobenzene	11.44	112	83189	3.75	ug/l	99
85) sec-Butylbenzene	13.17	105	22783	0.81	ug/l	88
89) n-Butylbenzene	13.62	91	12758	0.65	ug/l #	73

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN080418\
Data File : VN050366.D
Acq On : 4 Aug 2018 14:57
Operator : MD\SY
Sample : J4328-06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-03

Quant Time: Aug 10 09:02:47 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072418W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 26 18:10:10 2018
Response via : Initial Calibration

