

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN080418\
 Data File : VN050369.D
 Acq On : 4 Aug 2018 16:11
 Operator : MD\SY
 Sample : J4356-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-11

Quant Time: Aug 13 05:58:35 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 26 18:10:10 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	688989	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1086188	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	956918	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	409635	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	469742	48.87	ug/l	0.00
Spiked Amount						
			Recovery	=	97.74%	
35) Dibromofluoromethane	7.59	113	416834	47.00	ug/l	0.00
Spiked Amount						
			Recovery	=	94.00%	
50) Toluene-d8	10.09	98	1525827	46.80	ug/l	0.00
Spiked Amount						
			Recovery	=	93.60%	
62) 4-Bromofluorobenzene	12.40	95	462149	43.05	ug/l	0.00
Spiked Amount						
			Recovery	=	86.10%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
11) Tert butyl alcohol	4.80	59	44190	62.48	ug/l	97
16) Acetone	3.82	43	82400	32.61	ug/l	98
40) Benzene	8.04	78	13836	0.37	ug/l #	89
65) Chlorobenzene	11.44	112	15378	0.64	ug/l	96
95) Naphthalene	15.14	128	122627	10.39	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN080418\
Data File : VN050369.D
Acq On : 4 Aug 2018 16:11
Operator : MD\SY
Sample : J4356-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-11

Quant Time: Aug 13 05:58:35 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072418W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 26 18:10:10 2018
Response via : Initial Calibration

