

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN080418\
 Data File : VN050372.D
 Acq On : 4 Aug 2018 17:26
 Operator : MD\SY
 Sample : J4356-05
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-07

Quant Time: Aug 06 13:56:51 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 26 18:10:10 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	667100	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1048578	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	942966	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	411687	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	462769	49.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.44%	
35) Dibromofluoromethane	7.59	113	408767	47.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.50%	
50) Toluene-d8	10.09	98	1471537	46.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.50%	
62) 4-Bromofluorobenzene	12.40	95	457598	44.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.30%	
Target Compounds						
16) Acetone	3.82	43	33472	11.12	ug/l	99
19) Methyl tert-butyl Ether	5.05	73	602014	29.38	ug/l	100
40) Benzene	8.04	78	25563	0.71	ug/l	100
65) Chlorobenzene	11.44	112	234879	9.86	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN080418\
Data File : VN050372.D
Acq On : 4 Aug 2018 17:26
Operator : MD\SY
Sample : J4356-05
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-07

Quant Time: Aug 06 13:56:51 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072418W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 26 18:10:10 2018
Response via : Initial Calibration

