

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080422\
 Data File : VN073729.D
 Acq On : 04 Aug 2022 12:03
 Operator : JC\MD
 Sample : VN0804WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0804WBL01

Quant Time: Aug 05 00:24:32 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072822W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 29 02:20:25 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	372468	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	604275	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	569349	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.615	152	266911	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	222905	49.107	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	98.220%
35) Dibromofluoromethane	7.957	113	206045	54.301	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	108.600%
50) Toluene-d8	10.380	98	679859	45.531	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	91.060%
62) 4-Bromofluorobenzene	12.674	95	247042	49.598	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.200%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080422\
Data File : VN073729.D
Acq On : 04 Aug 2022 12:03
Operator : JC\MD
Sample : VN0804WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0804WBL01

Quant Time: Aug 05 00:24:32 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072822W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 29 02:20:25 2022
Response via : Initial Calibration

