

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080422\
 Data File : VN073733.D
 Acq On : 04 Aug 2022 13:51
 Operator : JC\MD
 Sample : N3887-07
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TB072522

Quant Time: Aug 05 00:27:08 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072822W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 29 02:20:25 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.022	168	375121	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	608619	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	572463	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.615	152	260633	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	230259	50.369	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 100.740%	
35) Dibromofluoromethane	7.957	113	210224	55.007	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 110.020%	
50) Toluene-d8	10.380	98	690818	45.935	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 91.880%	
62) 4-Bromofluorobenzene	12.674	95	251124	49.983	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 99.960%	
Target Compounds						Qvalue
16) Acetone	4.245	43	6119	2.492	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080422\
Data File : VN073733.D
Acq On : 04 Aug 2022 13:51
Operator : JC\MD
Sample : N3887-07
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TB072522

Quant Time: Aug 05 00:27:08 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072822W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 29 02:20:25 2022
Response via : Initial Calibration

