

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080422\
 Data File : VN073742.D
 Acq On : 04 Aug 2022 17:33
 Operator : JC\MD
 Sample : N4029-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 EFF0822

Quant Time: Aug 05 00:31:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072822W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 29 02:20:25 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	440952	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	684033	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	670892	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.615	152	368591	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.375	65	239157	44.505	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	89.000%
35) Dibromofluoromethane	7.957	113	228484	53.194	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	106.380%
50) Toluene-d8	10.380	98	762345	45.102	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	90.200%
62) 4-Bromofluorobenzene	12.674	95	313821	54.599	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	109.200%
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	5.287	59	32369	27.930	ug/l #	87
95) Naphthalene	15.433	128	84282	4.024	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080422\
Data File : VN073742.D
Acq On : 04 Aug 2022 17:33
Operator : JC\MD
Sample : N4029-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
EFF0822

Quant Time: Aug 05 00:31:17 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072822W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 29 02:20:25 2022
Response via : Initial Calibration

