

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080422\
 Data File : VN073744.D
 Acq On : 04 Aug 2022 18:23
 Operator : JC\MD
 Sample : PB146733TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB146733TB

Quant Time: Aug 05 00:32:04 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072822W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 29 02:20:25 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	417851	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	666013	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	637928	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.615	152	293850	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.375	65	235276	46.203	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	92.400%		
35) Dibromofluoromethane	7.951	113	221009	52.846	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	105.700%		
50) Toluene-d8	10.380	98	759275	46.136	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	92.280%		
62) 4-Bromofluorobenzene	12.674	95	287889	51.964	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	103.920%		
Target Compounds					Qvalue	
16) Acetone	4.228	43	76733	28.055	ug/l	99
20) Methylene Chloride	5.016	84	24620	3.515	ug/l	92
37) Ethyl Acetate	7.339	43	69571	8.863	ug/l	98
43) Isopropyl Acetate	8.498	43	276288	26.058	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080422\
Data File : VN073744.D
Acq On : 04 Aug 2022 18:23
Operator : JC\MD
Sample : PB146733TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB146733TB

Quant Time: Aug 05 00:32:04 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072822W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 29 02:20:25 2022
Response via : Initial Calibration

