

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080422\
 Data File : VN073745.D
 Acq On : 04 Aug 2022 18:48
 Operator : JC\MD
 Sample : N4023-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP1-1-6

Quant Time: Aug 05 00:32:30 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072822W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 29 02:20:25 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	425470	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	684659	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.687	117	670651	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.616	152	332359	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	240482	46.380	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	92.760%
35) Dibromofluoromethane	7.951	113	224896	52.311	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.620%
50) Toluene-d8	10.381	98	782564	46.256	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	92.520%
62) 4-Bromofluorobenzene	12.675	95	308006	53.718	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	107.440%
Target Compounds						
					Qvalue	
16) Acetone	4.228	43	75290	27.035	ug/l	95
18) Methyl Acetate	4.793	43	7094	1.024	ug/l #	83
20) Methylene Chloride	5.010	84	39858	6.949	ug/l	95
37) Ethyl Acetate	7.346	43	67682	8.387	ug/l	99
43) Isopropyl Acetate	8.498	43	248763	22.823	ug/l #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080422\
Data File : VN073745.D
Acq On : 04 Aug 2022 18:48
Operator : JC\MD
Sample : N4023-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP1-1-6

Quant Time: Aug 05 00:32:30 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072822W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 29 02:20:25 2022
Response via : Initial Calibration

