

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080422\
 Data File : VN073746.D
 Acq On : 04 Aug 2022 19:12
 Operator : JC\MD
 Sample : N4023-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP2-1-6

Quant Time: Aug 05 00:33:00 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072822W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 29 02:20:25 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.022	168	421201	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	682852	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	669552	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.616	152	328529	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	240662	46.885	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	93.780%
35) Dibromofluoromethane	7.951	113	222570	51.907	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.820%
50) Toluene-d8	10.380	98	787785	46.688	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	93.380%
62) 4-Bromofluorobenzene	12.674	95	305950	53.538	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	107.080%
Target Compounds						
					Qvalue	
16) Acetone	4.222	43	76554	27.767	ug/l	98
18) Methyl Acetate	4.781	43	8364	1.219	ug/l #	83
20) Methylene Chloride	5.016	84	35968	6.131	ug/l	98
37) Ethyl Acetate	7.339	43	64890	8.063	ug/l	99
43) Isopropyl Acetate	8.498	43	258144	23.747	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080422\
Data File : VN073746.D
Acq On : 04 Aug 2022 19:12
Operator : JC\MD
Sample : N4023-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP2-1-6

Quant Time: Aug 05 00:33:00 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072822W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 29 02:20:25 2022
Response via : Initial Calibration

