

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080422\
 Data File : VN073747.D
 Acq On : 04 Aug 2022 19:37
 Operator : JC\MD
 Sample : N4023-08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP3-1-6

Quant Time: Aug 05 00:33:28 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072822W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 29 02:20:25 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	417267	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	677486	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	652133	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.615	152	293218	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.375	65	239320	47.063	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	94.120%
35) Dibromofluoromethane	7.957	113	224754	52.831	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	105.660%
50) Toluene-d8	10.380	98	780071	46.597	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	93.200%
62) 4-Bromofluorobenzene	12.674	95	284423	50.713	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.420%
Target Compounds						
					Qvalue	
16) Acetone	4.228	43	75153	27.516	ug/l	97
18) Methyl Acetate	4.787	43	9272	1.365	ug/l	93
20) Methylene Chloride	5.010	84	32294	5.340	ug/l	95
37) Ethyl Acetate	7.339	43	65547	8.209	ug/l	99
43) Isopropyl Acetate	8.498	43	275930	25.584	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080422\
Data File : VN073747.D
Acq On : 04 Aug 2022 19:37
Operator : JC\MD
Sample : N4023-08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP3-1-6

Quant Time: Aug 05 00:33:28 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072822W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 29 02:20:25 2022
Response via : Initial Calibration

