

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN080519\
 Data File : VN057073.D
 Acq On : 5 Aug 2019 11:08
 Operator : JC/SP
 Sample : K4142-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 399L-241L

Quant Time: Aug 06 01:32:06 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072719W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jul 27 01:33:29 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	584594	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	833105	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	680334	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	184361	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	280253	46.83	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.66%	
35) Dibromofluoromethane	7.58	113	256658	48.12	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.24%	
50) Toluene-d8	10.08	98	970067	47.70	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.40%	
62) 4-Bromofluorobenzene	12.40	95	254063	38.37	ug/l	0.00
Spiked Amount	50.000		Recovery	=	76.74%	
Target Compounds						
27) cis-1,2-Dichloroethene	6.82	96	3165	0.490	ug/l	70
44) Trichloroethene	8.82	130	9146	1.470	ug/l	97
64) Tetrachloroethene	10.63	164	118355	20.188	ug/l	99
68) m/p-Xylenes	11.62	106	5853	0.662	ug/l	94
69) o-Xylene	11.94	106	2642	0.305	ug/l	98
95) Naphthalene	15.12	128	6128	4.314	ug/l #	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN080519\
Data File : VN057073.D
Acq On : 5 Aug 2019 11:08
Operator : JC/SP
Sample : K4142-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
399L-241L

Quant Time: Aug 06 01:32:06 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N072719W.M
Quant Title : SW846 8260
QLast Update : Sat Jul 27 01:33:29 2019
Response via : Initial Calibration

