

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN080519\
 Data File : VN057078.D
 Acq On : 5 Aug 2019 13:05
 Operator : JC/SP
 Sample : K4093-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PL-02-080119

Quant Time: Aug 06 02:09:37 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072719W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jul 27 01:33:29 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	562413	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	815440	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	685294	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	194163	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.02	65	279946	48.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.26%	
35) Dibromofluoromethane	7.58	113	248562	47.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.24%	
50) Toluene-d8	10.08	98	951058	47.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.56%	
62) 4-Bromofluorobenzene	12.40	95	260234	40.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	80.32%	
Target Compounds						
16) Acetone	3.81	43	21001	10.069	ug/l	97
20) Methylene Chloride	4.54	84	151688	25.248	ug/l	98
43) Isopropyl Acetate	8.16	43	21174	2.229	ug/l	91
95) Naphthalene	15.12	128	12752	5.023	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN080519\
Data File : VN057078.D
Acq On : 5 Aug 2019 13:05
Operator : JC/SP
Sample : K4093-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PL-02-080119

Quant Time: Aug 06 02:09:37 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N072719W.M
Quant Title : SW846 8260
QLast Update : Sat Jul 27 01:33:29 2019
Response via : Initial Calibration

