

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN080519\
 Data File : VN057083.D
 Acq On : 5 Aug 2019 15:38
 Operator : JC/SP
 Sample : K4135-32
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 1S-B1-(0-1)

Quant Time: Aug 06 02:19:00 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072719W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jul 27 01:33:29 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	568762	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	836787	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	684569	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	182142	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	283679	48.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.44%	
35) Dibromofluoromethane	7.58	113	255130	47.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.26%	
50) Toluene-d8	10.08	98	968725	47.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.84%	
62) 4-Bromofluorobenzene	12.40	95	250362	37.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	75.30%	
Target Compounds						
16) Acetone	3.82	43	15284	7.246	ug/l	92
20) Methylene Chloride	4.54	84	16059	2.643	ug/l	89
43) Isopropyl Acetate	8.17	43	26193	2.687	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN080519\
Data File : VN057083.D
Acq On : 5 Aug 2019 15:38
Operator : JC/SP
Sample : K4135-32
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
1S-B1-(0-1)

Quant Time: Aug 06 02:19:00 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072719W.M
Quant Title : SW846 8260
QLast Update : Sat Jul 27 01:33:29 2019
Response via : Initial Calibration

