

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN080519\
 Data File : VN057084.D
 Acq On : 5 Aug 2019 16:01
 Operator : JC/SP
 Sample : K4135-27
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 1S-A1-(30)

Quant Time: Aug 06 02:20:27 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072719W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jul 27 01:33:29 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	570062	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	833263	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	697066	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	192744	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.02	65	284201	48.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.40%	
35) Dibromofluoromethane	7.58	113	256278	48.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.08%	
50) Toluene-d8	10.08	98	970523	47.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.42%	
62) 4-Bromofluorobenzene	12.40	95	263174	39.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	79.48%	
Target Compounds						
8) Diethyl Ether	3.40	74	3442	1.086	ug/l	95
16) Acetone	3.81	43	14368	6.796	ug/l	99
20) Methylene Chloride	4.54	84	16010	2.629	ug/l	94
43) Isopropyl Acetate	8.17	43	23625	2.434	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN080519\
Data File : VN057084.D
Acq On : 5 Aug 2019 16:01
Operator : JC/SP
Sample : K4135-27
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
1S-A1-(30)

Quant Time: Aug 06 02:20:27 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072719W.M
Quant Title : SW846 8260
QLast Update : Sat Jul 27 01:33:29 2019
Response via : Initial Calibration

