

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN080519\
 Data File : VN057087.D
 Acq On : 5 Aug 2019 17:11
 Operator : JC/SP
 Sample : K4135-22
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 1S-A1-(0-1)

Quant Time: Aug 06 02:25:45 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072719W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jul 27 01:33:29 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	560493	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	822434	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	684722	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	182704	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.02	65	279371	48.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.38%	
35) Dibromofluoromethane	7.58	113	247689	47.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.08%	
50) Toluene-d8	10.09	98	960458	47.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.68%	
62) 4-Bromofluorobenzene	12.40	95	254330	38.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	77.82%	
Target Compounds						
8) Diethyl Ether	3.40	74	3454	1.108	ug/l	94
16) Acetone	3.82	43	18603	8.950	ug/l	97
18) Methyl Acetate	4.32	43	6872	1.036	ug/l #	80
20) Methylene Chloride	4.54	84	17812	2.975	ug/l	96
43) Isopropyl Acetate	8.17	43	22301	2.328	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN080519\
Data File : VN057087.D
Acq On : 5 Aug 2019 17:11
Operator : JC/SP
Sample : K4135-22
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
1S-A1-(0-1)

Quant Time: Aug 06 02:25:45 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N072719W.M
Quant Title : SW846 8260
QLast Update : Sat Jul 27 01:33:29 2019
Response via : Initial Calibration

