

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN080519\
 Data File : VN057075.D
 Acq On : 5 Aug 2019 11:55
 Operator : JC/SP
 Sample : PB121960TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB121960TB

Quant Time: Aug 06 01:36:08 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072719W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jul 27 01:33:29 2019
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	593046	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	864040	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	722725	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	196596	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.02	65	291093	47.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.90%	
35) Dibromofluoromethane	7.58	113	267509	48.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.72%	
50) Toluene-d8	10.08	98	1012695	48.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.02%	
62) 4-Bromofluorobenzene	12.40	95	271781	39.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	79.16%	
Target Compounds						
16) Acetone	3.81	43	12908	5.869	ug/l	99
20) Methylene Chloride	4.54	84	16888	2.666	ug/l	92
43) Isopropyl Acetate	8.16	43	18983	1.886	ug/l #	92
95) Naphthalene	15.12	128	5712	4.223	ug/l	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN080519\
Data File : VN057075.D
Acq On : 5 Aug 2019 11:55
Operator : JC/SP
Sample : PB121960TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB121960TB

Quant Time: Aug 06 01:36:08 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N072719W.M
Quant Title : SW846 8260
QLast Update : Sat Jul 27 01:33:29 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.66 min Scan# 1830

Delta R.T. 0.01 min

Lab File: VN057075.D

Acq: 5 Aug 2019 11:55

Instrument :

MSVOA_N

ClientSampleId :


PB121960TB

Tot Ion:168 Resp: 593046

Ion Ratio Lower Upper

168 100

99 48.8 39.0 58.4



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VNO

7.66

250000

200000



150000

100000

50000

0

Time-->



#16

Acetone

Concen: 5.869 ug/l

RT: 3.81 min Scan# 634

Delta R.T. 0.01 min

Lab File: VN057075.D


Acq: 5 Aug 2019 11:55

Tgt Ion: 43 Resp: 12908

Ion Ratio Lower Upper

43 100

58 34.5 27.2 40.8



Abundance Ion 43.00 (42.70 to 43.70): VNO

Ion 58.00 (57.70 to 58.70): VNO

3.81

5000

4000


3000


2000

1000

0

Time-->





#20

Methylene Chloride

Concen: 2.666 ug/l

RT: 4.54 min Scan# 860

Delta R.T. 0.01 min

Lab File: VN057075.D

Acq: 5 Aug 2019 11:55

Instrument :

MSVOA_N

ClientSampled :

PB121960TB

Tot Ion: 84 Resp: 16888


Ion Ratio Lower Upper

84 100

49 105.3 95.2 142.8

51 40.3 29.7 44.5

86 65.0 51.0 76.6






Abundance Ion 83.90 (83.60 to 84.60): VN0

Ion 48.90 (48.60 to 49.60): VN0

Ion 50.90 (50.60 to 51.60): VN0

Ion 85.90 (85.60 to 86.60): VN0



#33

1,2-Dichloroethane-d4

Concen: 47.951 ug/l

RT: 8.02 min Scan# 1942

Delta R.T. 0.01 min

Lab File: VN057075.D


Acq: 5 Aug 2019 11:55

Tgt Ion: 65 Resp: 291093

Ion Ratio Lower Upper



65 100


67 54.6 0.0 112.0



Abundance Ion 64.90 (64.60 to 65.60): VN0

Ion 66.90 (66.60 to 67.60): VN0









#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 8.58 min Scan# 2116
 Delta R.T. 0.00 min
 Lab File: VN057075.D
 Acq: 5 Aug 2019 11:55

Instrument :
 MSVOA_N
 ClientSampled :
 PB121960TB

Tot Ion:114 Resp: 864040
 Ion Ratio Lower Upper
 114 100
 63 19.0 0.0 38.0
 88 15.1 0.0 30.0




Abundance Ion 113.90 (113.60 to 114.60): V
 Ion 63.00 (62.70 to 63.70): VN
 Ion 87.90 (87.60 to 88.60): VN






#35
 Dibromofluoromethane
 Concen: 48.363 ug/l
 RT: 7.58 min Scan# 1806
 Delta R.T. 0.01 min
 Lab File: VN057075.D
 Acq: 5 Aug 2019 11:55

Tgt Ion:113 Resp: 267509
 Ion Ratio Lower Upper
 113 100
 111 101.0 82.0 123.0
 192 23.5 18.4 27.6



Abundance Ion 112.80 (112.50 to 113.50): V
 Ion 110.80 (110.50 to 111.50): V
 Ion 191.70 (191.40 to 192.40): V





#43

Isopropyl Acetate

Concen: 1.886 ug/l

RT: 8.16 min Scan# 1985

Delta R.T. 0.01 min

Lab File: VN057075.D

Acq: 5 Aug 2019 11:55

Instrument :

MSVOA_N

ClientSampleId :

PB121960TB


Tot Ion: 43 Resp: 18983

Ion Ratio Lower Upper

43 100

61 22.2 23.0 34.6#




87 14.2 11.7 17.5



Abundance Ion 43.00 (42.70 to 43.70): VN056941.D (-1962) (-)

Ion 61.00 (60.70 to 61.70): VN056941.D (-1962) (-)

Ion 87.00 (86.70 to 87.70): VN056941.D (-1962) (-)



#50

Toluene-d8

Concen: 48.010 ug/l

RT: 10.08 min Scan# 2584

Delta R.T. 0.00 min

Lab File: VN057075.D


Acq: 5 Aug 2019 11:55

Tgt Ion: 98 Resp: 1012695

Ion Ratio Lower Upper



98 100


100 64.9 51.3 76.9



Abundance Ion 98.00 (97.70 to 98.70): VN056941.D (-2566) (-)

Ion 100.00 (99.70 to 100.70): VN056941.D (-2566) (-)





#62

4-Bromofluorobenzene

Concen: 39.579 ug/l

RT: 12.40 min Scan# 3305

Delta R.T. 0.01 min

Lab File: VN057075.D

Acq: 5 Aug 2019 11:55

Instrument :

MSVOA_N

ClientSampleId :

PB121960TB


Tot Ion: 95 Resp: 271781

Ion Ratio Lower Upper

95 100

174 94.2 0.0 182.2

176 90.4 0.0 179.4



Abundance Ion 94.90 (94.60 to 95.60): VN0

Ion 173.80 (173.50 to 174.50): VN0

Ion 175.80 (175.50 to 176.50): VN0

12.40

200000



150000

100000

50000

0

Time--> 12.30 12.40 12.50



#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.40 min Scan# 2995

Delta R.T. 0.00 min

Lab File: VN057075.D

Acq: 5 Aug 2019 11:55


Tgt Ion: 117 Resp: 722725

Ion Ratio Lower Upper

117 100

82 50.9 42.2 63.4

119 31.2 25.8 38.6



Abundance Ion 116.90 (116.60 to 117.60): VN0

Ion 82.00 (81.70 to 82.70): VN0

Ion 118.90 (118.60 to 119.60): VN0

11.40

500000

400000


300000


200000

100000

0

Time--> 11.30 11.40 11.50









#72
 1,4-Dichlorobenzene-d4
 Concen: 50.000 ug/l
 RT: 13.34 min Scan# 3598
 Delta R.T. 0.00 min
 Lab File: VN057075.D
 Acq: 5 Aug 2019 11:55

Instrument :
 MSVOA_N
 ClientSampleId :
 PB121960TB

Tot Ion:152 Resp: 196596
 Ion Ratio Lower Upper
 152 100
 115 54.9 28.2 84.6
 150 154.2 0.0 349.4




Abundance Ion 151.90 (151.60 to 152.60): V
 Ion 114.90 (114.60 to 115.60): V
 Ion 149.90 (149.60 to 150.60): V



#95
 Naphthalene
 Concen: 4.223 ug/l
 RT: 15.12 min Scan# 4152
 Delta R.T. 0.00 min
 Lab File: VN057075.D
 Acq: 5 Aug 2019 11:55

Tgt Ion:128 Resp: 5712
 Ion Ratio Lower Upper
 128 100
 127 11.4 10.0 15.0
 129 9.5 8.7 13.1



Abundance Ion 128.00 (127.70 to 128.70): V
 Ion 127.00 (126.70 to 127.70): V
 Ion 129.00 (128.70 to 129.70): V

