

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080521\
 Data File : VN068017.D
 Acq On : 5 Aug 2021 15:43
 Operator : JC/MD
 Sample : PB138172TB
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB138172TB

Manual Integrations
 APPROVED

MMDadoda
 8/6/2021 5:42:59 PM

Quant Time: Aug 06 01:23:20 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072921W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 29 15:02:35 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	306703	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.971	114	594110	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	620174	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	216196	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	236510	54.471	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery	= 108.940%		
35) Dibromofluoromethane	8.024	113	180211	52.048	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery	= 104.100%		
50) Toluene-d8	10.443	98	789602	49.460	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery	= 98.920%		
62) 4-Bromofluorobenzene	12.734	95	261938	45.304	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery	= 90.600%		
Target Compounds						Qvalue
16) Acetone	4.296	43	37577	23.719	ug/l	94
20) Methylene Chloride	5.101	84	31145m	8.277	ug/l	
43) Isopropyl Acetate	8.568	43	37288	4.327	ug/l	# 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080521\
Data File : VN068017.D
Acq On : 5 Aug 2021 15:43
Operator : JC/MD
Sample : PB138172TB
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampled :
PB138172TB

Manual Integrations
APPROVED

MMDadoda
8/6/2021 5:42:59 PM

Quant Time: Aug 06 01:23:20 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072921W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 29 15:02:35 2021
Response via : Initial Calibration

