

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080521\
 Data File : VN068020.D
 Acq On : 5 Aug 2021 16:56
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Manual Integrations
 APPROVED

MMDadoda
 8/6/2021 5:42:57 PM

Quant Time: Aug 06 01:24:06 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072921W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 29 15:02:35 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	297636	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	585071	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	564299	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	187384	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	230958	54.813	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 109.620%	
35) Dibromofluoromethane	8.021	113	178641	52.392	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 104.780%	
50) Toluene-d8	10.443	98	744632	47.527	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 95.060%	
62) 4-Bromofluorobenzene	12.734	95	237553	42.115	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 84.220%	
Target Compounds						
20) Methylene Chloride	5.079	84	7397m	1.810	ug/l	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080521\
Data File : VN068020.D
Acq On : 5 Aug 2021 16:56
Operator : JC/MD
Sample : IBLK
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Manual Integrations
APPROVED

MMDadoda
8/6/2021 5:42:57 PM

Quant Time: Aug 06 01:24:06 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072921W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 29 15:02:35 2021
Response via : Initial Calibration

