

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080521\
 Data File : VN068030.D
 Acq On : 5 Aug 2021 21:00
 Operator : JC/MD
 Sample : M3291-01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-1

Quant Time: Aug 06 01:26:59 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072921W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 29 15:02:35 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	298075	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	584729	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	621448	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	212653	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	234359	55.538	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	111.080%
35) Dibromofluoromethane	8.021	113	179448	52.659	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	105.320%
50) Toluene-d8	10.446	98	779895	49.622	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	99.240%
62) 4-Bromofluorobenzene	12.733	95	258359	45.391	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	90.780%
Target Compounds						
					Qvalue	
16) Acetone	4.293	43	27804	18.058	ug/l #	89
18) Methyl Acetate	4.862	43	9167	2.379	ug/l #	75
20) Methylene Chloride	5.095	84	32509	8.910	ug/l	92
25) 2-Butanone	7.348	43	20657	9.248	ug/l #	89
43) Isopropyl Acetate	8.568	43	41069	4.842	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080521\
Data File : VN068030.D
Acq On : 5 Aug 2021 21:00
Operator : JC/MD
Sample : M3291-01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-1

Quant Time: Aug 06 01:26:59 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072921W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 29 15:02:35 2021
Response via : Initial Calibration

