

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080522\
 Data File : VN073763.D
 Acq On : 05 Aug 2022 15:30
 Operator : JC\MD
 Sample : N3887-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-12

Quant Time: Aug 07 20:51:24 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072822W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 29 02:20:25 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.022	168	404253	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	620196	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	600387	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.615	152	318993	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.375	65	220337	44.725	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	89.460%
35) Dibromofluoromethane	7.957	113	210746	54.115	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	108.220%
50) Toluene-d8	10.380	98	729569	47.606	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.220%
62) 4-Bromofluorobenzene	12.674	95	279822	53.849	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	107.700%
Target Compounds						
					Qvalue	
16) Acetone	4.234	43	9109	3.442	ug/l	92
31) Cyclohexane	8.022	56	157297	21.446	ug/l	98
39) Methylcyclohexane	9.398	83	97901	14.760	ug/l	99
40) Benzene	8.386	78	22681	1.258	ug/l	92
68) m/p-Xylenes	11.898	106	6700	0.817	ug/l	81
69) o-Xylene	12.216	106	3641	0.455	ug/l	74
83) tert-Butylbenzene	13.263	119	11783	0.692	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080522\
Data File : VN073763.D
Acq On : 05 Aug 2022 15:30
Operator : JC\MD
Sample : N3887-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-12

Quant Time: Aug 07 20:51:24 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072822W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 29 02:20:25 2022
Response via : Initial Calibration

