

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080525\
 Data File : VN087451.D
 Acq On : 05 Aug 2025 12:58
 Operator : JC\MD
 Sample : VN0805WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0805WBL01

Quant Time: Aug 06 02:11:30 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N080525W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Aug 06 02:01:21 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.800	128	70384	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	9.082	114	354106	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.847	117	314034	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.565	65	176744	27.225	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	90.767%#
60) 4-Bromofluorobenzene	12.829	95	151738	28.077	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	93.600%
63) Toluene-d8	10.547	98	431667	29.874	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	99.567%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080525\
Data File : VN087451.D
Acq On : 05 Aug 2025 12:58
Operator : JC\MD
Sample : VN0805WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0805WBL01

Quant Time: Aug 06 02:11:30 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N080525W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Wed Aug 06 02:01:21 2025
Response via : Initial Calibration

