

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN080618\
 Data File : VN050392.D
 Acq On : 6 Aug 2018 18:45
 Operator : MD\SY
 Sample : PB11760ZHE#07
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB11760ZHE#07

Quant Time: Aug 07 05:49:51 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080618W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 07 05:22:15 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	668143	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1108629	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	974997	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	381737	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	473536	53.05	ug/l	0.00
Spiked Amount			Recovery	=	106.10%	
35) Dibromofluoromethane	7.59	113	418314	47.03	ug/l	0.00
Spiked Amount			Recovery	=	94.06%	
50) Toluene-d8	10.09	98	1520370	45.46	ug/l	0.00
Spiked Amount			Recovery	=	90.92%	
62) 4-Bromofluorobenzene	12.40	95	443837	38.94	ug/l	0.00
Spiked Amount			Recovery	=	77.88%	


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
3) Chloromethane	2.06	50	4186	Below Cal		90
5) Bromomethane	2.56	94	4190	1.07 ug/l		88
10) Methyl Iodide	3.96	142	1829	0.54 ug/l #		65
13) Acrolein	3.62	56	151	15.18 ug/l #		15
16) Acetone	3.82	43	20923	7.51 ug/l		96
18) Methyl Acetate	4.33	43	16096	2.14 ug/l		99
20) Methylene Chloride	4.55	84	14511	0.24 ug/l		96
25) 2-Butanone	6.84	43	1507	0.34 ug/l		93
29) Tetrahydrofuran	7.23	42	1663	0.70 ug/l #		56
31) Cyclohexane	7.67	56	11646	0.28 ug/l #		1
36) 1,1-Dichloropropene	7.67	75	47002	3.86 ug/l #		49
38) Carbon Tetrachloride	7.67	117	59130	4.65 ug/l #		16
43) Isopropyl Acetate	8.17	43	452677	29.89 ug/l #		92
48) Methyl methacrylate	9.18	41	10415	1.37 ug/l #		23
49) 1,4-Dioxane	9.21	88	73	0.62 ug/l #		22
51) 4-Methyl-2-Pentanone	9.99	43	2994	0.27 ug/l		88
54) cis-1,3-Dichloropropene	9.91	75	3056	0.22 ug/l #		59
56) Ethyl methacrylate	10.64	69	1004	1.79 ug/l #		1
59) 2-Hexanone	10.77	43	26633	3.50 ug/l #		58
76) 1,2,3-Trichloropropane	12.40	75	211433	30.60 ug/l #		100
80) 1,3,5-Trimethylbenzene	12.73	105	6075	0.26 ug/l		94
81) trans-1,4-Dichloro-2-buten	12.40	75	211433	107.51 ug/l #		13
92) 1,2-Dibromo-3-Chloropropan	14.26	75	301	0.21 ug/l #		1
95) Naphthalene	15.14	128	768	3.66 ug/l #		69


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN080618\
Data File : VN050392.D
Acq On : 6 Aug 2018 18:45
Operator : MD\SY
Sample : PB11760ZHE#07
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB11760ZHE#07

Quant Time: Aug 07 05:49:51 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N080618W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 07 05:22:15 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.00 ug/l

RT: 7.67 min Scan# 1891

Delta R.T. 0.00 min

Lab File: VN050392.D

Acq: 6 Aug 2018 18:45

Instrument :

MSVOA_N

ClientSampleId :


PB11760ZHE#07

Tot Ion:168 Resp: 668143

Ion Ratio Lower Upper

168 100

99 50.6 40.0 60.0



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VNO

7.67

300000

250000

200000



150000

100000

50000

0

Time-->



#3

Chloromethane

Concen: Below Cal

RT: 2.06 min Scan# 146

Delta R.T. 0.00 min

Lab File: VN050392.D


Acq: 6 Aug 2018 18:45

Tgt Ion: 50 Resp: 4186

Ion Ratio Lower Upper

50 100

52 37.8 25.9 38.9



Abundance Ion 49.90 (49.60 to 50.60): VNO

Ion 51.90 (51.60 to 52.60): VNO

3000

2500

2000


1500


1000

500

0

Time-->





#5

Bromomethane

Concen: 1.07 ug/l

RT: 2.56 min Scan# 303

Delta R.T. 0.01 min

Lab File: VN050392.D

Acq: 6 Aug 2018 18:45

Instrument :

MSVOA_N

ClientSampleId :


PB11760ZHE#07

Tot Ion: 94 Resp: 4190

Ion Ratio Lower Upper




94 100

96 81.7 74.6 112.0



Abundance Ion 93.90 (93.60 to 94.60): VN050378.D (-284) (-)

Ion 95.90 (95.60 to 96.60): VN050392.D (-207) (-)



#10

Methyl Iodide

Concen: 0.54 ug/l

RT: 3.96 min Scan# 737

Delta R.T. 0.01 min

Lab File: VN050392.D

Acq: 6 Aug 2018 18:45


Tgt Ion: 142 Resp: 1829

Ion Ratio Lower Upper

142 100

127 20.1 33.4 50.2#



141 0.0 11.6 17.4#




Abundance Ion 141.80 (141.50 to 142.50): VN050378.D (-713) (-)

Ion 126.80 (126.50 to 127.50): VN050392.D (-639) (-)

Ion 140.80 (140.50 to 141.50): VN050392.D (-639) (-)





#13

Acrolein

Concen: 15.18 ug/l

RT: 3.62 min Scan# 630

Delta R.T. 0.01 min

Lab File: VN050392.D

Acq: 6 Aug 2018 18:45

Instrument :

MSVOA_N

ClientSampleId :





PB11760ZHE#07

Tot Ion: 56 Resp: 151

Ion Ratio Lower Upper

56 100

55 0.0 55.8 83.6#



#16

Acetone

Concen: 7.51 ug/l

RT: 3.82 min Scan# 695

Delta R.T. 0.00 min

Lab File: VN050392.D




Acq: 6 Aug 2018 18:45


Tgt Ion: 43 Resp: 20923

Ion Ratio Lower Upper

43 100

58 34.5 25.8 38.8









#18
Methyl Acetate
Concen: 2.14 ug/l
RT: 4.33 min Scan# 853
Delta R.T. 0.00 min
Lab File: VN050392.D
Acq: 6 Aug 2018 18:45




Instrument :
MSVOA_N
ClientSampleId :
PB11760ZHE#07


Tot Ion: 43 Resp: 16096
Ion Ratio Lower Upper
43 100
74 24.6 19.2 28.8



#20
Methylene Chloride
Concen: 0.24 ug/l
RT: 4.55 min Scan# 921
Delta R.T. 0.00 min
Lab File: VN050392.D
Acq: 6 Aug 2018 18:45

Tgt Ion: 84 Resp: 14511
Ion Ratio Lower Upper
84 100
49 111.9 94.8 142.2
51 34.1 28.3 42.5
86 62.1 51.0 76.6





#25

2-Butanone

Concen: 0.34 ug/l

RT: 6.84 min Scan# 1632

Delta R.T. -0.00 min

Lab File: VN050392.D

Acq: 6 Aug 2018 18:45

Instrument :

MSVOA_N

ClientSampleId :





PB11760ZHE#07

Tot Ion: 43 Resp: 1507

Ion Ratio Lower Upper

43 100

72 30.3 21.5 32.3



#29

Tetrahydrofuran

Concen: 0.70 ug/l

RT: 7.23 min Scan# 1754

Delta R.T. 0.01 min

Lab File: VN050392.D

Acq: 6 Aug 2018 18:45




Tgt Ion: 42 Resp: 1663


Ion Ratio Lower Upper

42 100

72 25.4 35.9 53.9#

71 4.1 33.7 50.5#









#31
Cyclohexane
Concen: 0.28 ug/l
RT: 7.67 min Scan# 1891
Delta R.T. 0.01 min
Lab File: VN050392.D
Acq: 6 Aug 2018 18:45

Instrument :
MSVOA_N
ClientSampleId :
PB11760ZHE#07

Tot Ion: 56 Resp: 11646
Ion Ratio Lower Upper
56 100
69 195.4 26.2 39.4#
84 159.7 67.4 101.0#




Abundance Ion 56.00 (55.70 to 56.70): VN0
Ion 69.00 (68.70 to 69.70): VN0
Ion 84.00 (83.70 to 84.70): VN0






#33
1,2-Dichloroethane-d4
Concen: 53.05 ug/l
RT: 8.03 min Scan# 2003
Delta R.T. -0.00 min
Lab File: VN050392.D
Acq: 6 Aug 2018 18:45

Tgt Ion: 65 Resp: 473536
Ion Ratio Lower Upper
65 100
67 53.5 0.0 108.4



Abundance Ion 64.90 (64.60 to 65.60): VN0
Ion 66.90 (66.60 to 67.60): VN0









#34
 1,4-Difluorobenzene
 Concen: 50.00 ug/l
 RT: 8.59 min Scan# 2176
 Delta R.T. -0.00 min
 Lab File: VN050392.D
 Acq: 6 Aug 2018 18:45

Instrument :
 MSVOA_N
 ClientSampleId :
 PB11760ZHE#07

Tot Ion:114 Resp: 1108629
 Ion Ratio Lower Upper
 114 100
 63 19.9 0.0 40.0
 88 15.4 0.0 31.2




Abundance Ion 113.90 (113.60 to 114.60): V
 Ion 63.00 (62.70 to 63.70): VN
 Ion 87.90 (87.60 to 88.60): VN






#35
 Dibromofluoromethane
 Concen: 47.03 ug/l
 RT: 7.59 min Scan# 1867
 Delta R.T. -0.00 min
 Lab File: VN050392.D
 Acq: 6 Aug 2018 18:45

Tgt Ion:113 Resp: 418314
 Ion Ratio Lower Upper
 113 100
 111 102.5 80.7 121.1
 192 21.9 17.8 26.6



Abundance Ion 112.80 (112.50 to 113.50): V
 Ion 110.80 (110.50 to 111.50): V
 Ion 191.70 (191.40 to 192.40): V





#36

1,1-Dichloropropene

Concen: 3.86 ug/l

RT: 7.67 min Scan# 1890

Delta R.T. -0.13 min

Lab File: VN050392.D

Acq: 6 Aug 2018 18:45

Instrument :

MSVOA_N

ClientSampleId :

PB11760ZHE#07





Tot Ion: 75 Resp: 47002

Ion Ratio Lower Upper

75 100

110 9.0 18.4 55.2#

77 0.1 25.3 37.9#



#38

Carbon Tetrachloride

Concen: 4.65 ug/l

RT: 7.67 min Scan# 1890

Delta R.T. -0.11 min

Lab File: VN050392.D

Acq: 6 Aug 2018 18:45




Tgt Ion: 117 Resp: 59130


Ion Ratio Lower Upper

117 100

119 5.6 77.4 116.0#

121 0.0 23.9 35.9#






#43




Isopropyl Acetate
 Concen: 29.89 ug/l
 RT: 8.17 min Scan# 2047
 Delta R.T. 0.01 min
 Lab File: VN050392.D
 Acq: 6 Aug 2018 18:45

Instrument :
 MSVOA_N
 ClientSampleId :
 PB11760ZHE#07

Tot Ion: 43 Resp: 452677
 Ion Ratio Lower Upper
 43 100
 61 15.6 15.8 23.6#
 87 10.0 10.3 15.5#




Abundance Ion 43.00 (42.70 to 43.70): VNC
 Ion 61.00 (60.70 to 61.70): VNC
 Ion 87.00 (86.70 to 87.70): VNC





#48


Methyl methacrylate
 Concen: 1.37 ug/l
 RT: 9.18 min Scan# 2362
 Delta R.T. -0.02 min
 Lab File: VN050392.D
 Acq: 6 Aug 2018 18:45

Tgt Ion: 41 Resp: 10415
 Ion Ratio Lower Upper
 41 100
 69 0.0 70.6 105.8#
 39 19.4 43.9 65.9#



Abundance Ion 41.00 (40.70 to 41.70): VNC
 Ion 69.00 (68.70 to 69.70): VNC
 Ion 39.00 (38.70 to 39.70): VNC





#49

1,4-Dioxane

Concen: 0.62 ug/l

RT: 9.21 min Scan# 2371

Delta R.T. 0.01 min

Lab File: VN050392.D

Acq: 6 Aug 2018 18:45

Instrument :

MSVOA_N

ClientSampleId :

PB11760ZHE#07


Tot Ion: 88 Resp: 73

Ion Ratio Lower Upper

88 100

43 0.0 26.9 40.3#




58 0.0 57.3 85.9#



Abundance Ion 88.00 (87.70 to 88.70): VN0

Ion 43.00 (42.70 to 43.70): VN0

Ion 58.00 (57.70 to 58.70): VN0



#50

Toluene-d8

Concen: 45.46 ug/l

RT: 10.09 min Scan# 2645

Delta R.T. 0.00 min

Lab File: VN050392.D


Acq: 6 Aug 2018 18:45

Tgt Ion: 98 Resp: 1520370

Ion Ratio Lower Upper



98 100


100 64.0 50.7 76.1



Abundance Ion 98.00 (97.70 to 98.70): VN0

Ion 100.00 (99.70 to 100.70): VN0









#51
 4-Methyl-2-Pentanone
 Concen: 0.27 ug/l
 RT: 9.99 min Scan# 2614
 Delta R.T. 0.01 min
 Lab File: VN050392.D
 Acq: 6 Aug 2018 18:45




Instrument :
 MSVOA_N
 ClientSampleId :
 PB11760ZHE#07


Tot Ion: 43 Resp: 2994
 Ion Ratio Lower Upper
 43 100
 58 32.4 31.9 47.9



#54
 cis-1,3-Dichloropropene
 Concen: 0.22 ug/l
 RT: 9.91 min Scan# 2587
 Delta R.T. 0.07 min
 Lab File: VN050392.D
 Acq: 6 Aug 2018 18:45

Tgt Ion: 75 Resp: 3056
 Ion Ratio Lower Upper
 75 100
 77 0.0 25.5 38.3#
 39 26.8 36.6 55.0#





#56

Ethyl methacrylate

Concen: 1.79 ug/l

RT: 10.64 min Scan# 2816

Delta R.T. 0.21 min

Lab File: VN050392.D

Acq: 6 Aug 2018 18:45

Instrument :

MSVOA_N

ClientSampleId :

PB11760ZHE#07


Tot Ion: 69 Resp: 1004

Ion Ratio Lower Upper

69 100

41 608.2 51.0 76.4#




39 339.4 26.6 40.0#



Abundance Ion 69.00 (68.70 to 69.70): VN050392.D (-2657) (-)

Ion 41.00 (40.70 to 41.70): VN050392.D (-2657) (-)

Ion 39.00 (38.70 to 39.70): VN050392.D (-2657) (-)



#59

2-Hexanone

Concen: 3.50 ug/l

RT: 10.77 min Scan# 2854

Delta R.T. 0.01 min

Lab File: VN050392.D


Acq: 6 Aug 2018 18:45

Tgt Ion: 43 Resp: 26633

Ion Ratio Lower Upper



43 100


58 25.0 27.6 82.7#



Abundance Ion 43.00 (42.70 to 43.70): VN050392.D (-2757) (-)

Ion 58.00 (57.70 to 58.70): VN050392.D (-2757) (-)









#62
4-Bromofluorobenzene
Concen: 38.94 ug/l
RT: 12.40 min Scan# 3363
Delta R.T. -0.00 min
Lab File: VN050392.D
Acq: 6 Aug 2018 18:45




Instrument :
MSVOA_N
ClientSampleId :
PB11760ZHE#07


Tot Ion: 95 Resp: 443837
Ion Ratio Lower Upper
95 100
174 90.7 0.0 178.2
176 86.4 0.0 173.4



#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 11.41 min Scan# 3054
Delta R.T. -0.00 min
Lab File: VN050392.D
Acq: 6 Aug 2018 18:45

Tgt Ion: 117 Resp: 974997
Ion Ratio Lower Upper
117 100
82 53.9 43.7 65.5
119 32.6 26.1 39.1





#72

1,4-Dichlorobenzene-d4

Concen: 50.00 ug/l

RT: 13.35 min Scan# 3656

Delta R.T. -0.00 min

Lab File: VN050392.D

Acq: 6 Aug 2018 18:45

Instrument :

MSVOA_N

ClientSampleId :

PB11760ZHE#07


Tot Ion:152 Resp: 381737

Ion Ratio Lower Upper

152 100

115 56.3 27.8 83.3




150 157.2 0.0 348.6



Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V



#76

1,2,3-Trichloropropane

Concen: 30.60 ug/l

RT: 12.40 min Scan# 3363

Delta R.T. -0.15 min

Lab File: VN050392.D


Acq: 6 Aug 2018 18:45

Tgt Ion: 75 Resp: 211433

Ion Ratio Lower Upper



75 100


77 1.2 0.0 0.0#



Abundance Ion 74.90 (74.60 to 75.60): V

Ion 77.00 (76.70 to 77.70): V









#80
 1,3,5-Trimethylbenzene
 Concen: 0.26 ug/l
 RT: 12.73 min Scan# 3466
 Delta R.T. -0.00 min
 Lab File: VN050392.D
 Acq: 6 Aug 2018 18:45




Instrument :
 MSVOA_N
 ClientSampleId :
 PB11760ZHE#07


Tot Ion:105 Resp: 6075
 Ion Ratio Lower Upper
 105 100
 120 44.7 24.4 73.2



#81
 trans-1,4-Dichloro-2-butene
 Concen: 107.51 ug/l
 RT: 12.40 min Scan# 3363
 Delta R.T. 0.10 min
 Lab File: VN050392.D
 Acq: 6 Aug 2018 18:45

Tgt Ion: 75 Resp: 211433
 Ion Ratio Lower Upper
 75 100
 53 0.2 73.5 110.3#
 89 0.0 36.0 54.0#





#92

1,2-Dibromo-3-Chloropropane

Concen: 0.21 ug/l

RT: 14.26 min Scan# 3939

Delta R.T. -0.02 min

Lab File: VN050392.D

Acq: 6 Aug 2018 18:45

Instrument :

MSVOA_N

ClientSampleId :

PB11760ZHE#07


Tot Ion: 75 Resp: 301

Ion Ratio Lower Upper

75 100

155 0.0 44.8 134.4#




157 0.0 57.2 171.6#



Abundance Ion 74.90 (74.60 to 75.60): VN0

Ion 154.80 (154.50 to 155.50): V

Ion 156.80 (156.50 to 157.50): V



#95

Naphthalene

Concen: 3.66 ug/l

RT: 15.14 min Scan# 4214

Delta R.T. 0.00 min

Lab File: VN050392.D

Acq: 6 Aug 2018 18:45


Tgt Ion: 128 Resp: 768

Ion Ratio Lower Upper

128 100

127 0.0 10.4 15.6#

129 0.0 8.8 13.2#



Abundance Ion 128.00 (127.70 to 128.70): V

Ion 127.00 (126.70 to 127.70): V

Ion 129.00 (128.70 to 129.70): V

