

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN080618\
 Data File : VN050392.D
 Acq On : 6 Aug 2018 18:45
 Operator : MD\SY
 Sample : PB11760ZHE#07
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB11760ZHE#07

Quant Time: Aug 07 06:47:53 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080618W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 07 05:22:15 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	668143	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1108629	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	974997	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	381737	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	473536	53.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.10%	
35) Dibromofluoromethane	7.59	113	418314	47.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.06%	
50) Toluene-d8	10.09	98	1520370	45.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.92%	
62) 4-Bromofluorobenzene	12.40	95	443837	38.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	77.88%	
Target Compounds						
16) Acetone	3.82	43	20923	7.51	ug/l	96
18) Methyl Acetate	4.33	43	16096	2.14	ug/l	99
43) Isopropyl Acetate	8.17	43	452677	29.89	ug/l #	92
59) 2-Hexanone	10.77	43	26633	3.50	ug/l #	58

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN080618\
Data File : VN050392.D
Acq On : 6 Aug 2018 18:45
Operator : MD\SY
Sample : PB11760ZHE#07
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB11760ZHE#07

Quant Time: Aug 07 06:47:53 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N080618W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 07 05:22:15 2018
Response via : Initial Calibration

