

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN080618\
 Data File : VN050396.D
 Acq On : 6 Aug 2018 20:25
 Operator : MD\SY
 Sample : PB11760ZHE#11
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB11760ZHE#11

Quant Time: Aug 07 06:57:54 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080618W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 07 05:22:15 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	657134	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1114808	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	948454	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	361039	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	472782	53.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.70%	
35) Dibromofluoromethane	7.59	113	419411	46.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.78%	
50) Toluene-d8	10.09	98	1511713	44.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.90%	
62) 4-Bromofluorobenzene	12.40	95	423735	36.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	73.94%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.82	43	22286	8.16	ug/l	96
18) Methyl Acetate	4.33	43	17461	2.38	ug/l	96
43) Isopropyl Acetate	8.17	43	428222	28.12	ug/l #	92
59) 2-Hexanone	10.77	43	25304	3.31	ug/l #	57

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN080618\
Data File : VN050396.D
Acq On : 6 Aug 2018 20:25
Operator : MD\SY
Sample : PB11760ZHE#11
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB11760ZHE#11

Quant Time: Aug 07 06:57:54 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N080618W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 07 05:22:15 2018
Response via : Initial Calibration

