

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN080618\
 Data File : VN050404.D
 Acq On : 7 Aug 2018 00:33
 Operator : MD\SY
 Sample : J4277-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 OR-02-080318-B

Quant Time: Aug 07 07:38:54 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080618W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 07 05:22:15 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	656492	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1081154	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	940593	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	380384	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	468199	53.38	ug/l	0.00
Spiked Amount						
			Recovery	=	106.76%	
35) Dibromofluoromethane	7.60	113	412197	47.52	ug/l	0.00
Spiked Amount						
			Recovery	=	95.04%	
50) Toluene-d8	10.09	98	1476031	45.25	ug/l	0.00
Spiked Amount						
			Recovery	=	90.50%	
62) 4-Bromofluorobenzene	12.40	95	430731	38.75	ug/l	0.00
Spiked Amount						
			Recovery	=	77.50%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.82	43	55411	20.85	ug/l	98
18) Methyl Acetate	4.34	43	14333	1.92	ug/l	93
20) Methylene Chloride	4.55	84	519884	65.47	ug/l	99
30) Chloroform	7.38	83	33513	2.35	ug/l	97
43) Isopropyl Acetate	8.17	43	215880	14.62	ug/l	99
59) 2-Hexanone	10.77	43	15649	2.11	ug/l #	60
95) Naphthalene	15.14	128	80647	7.68	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN080618\
Data File : VN050404.D
Acq On : 7 Aug 2018 00:33
Operator : MD\SY
Sample : J4277-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
OR-02-080318-B

Quant Time: Aug 07 07:38:54 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080618W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 07 05:22:15 2018
Response via : Initial Calibration

