

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN080618\
 Data File : VN050406.D
 Acq On : 7 Aug 2018 1:23
 Operator : MD\SY
 Sample : J4316-09
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 OW-2-4

Manual Integrations
 APPROVED

MMDadoda
 8/7/2018 10:46:54 AM

Quant Time: Aug 07 09:47:39 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080618W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 07 05:22:15 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	645477	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1031912	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	865649	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	334531	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	453690	52.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.22%	
35) Dibromofluoromethane	7.59	113	400831	48.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.82%	
50) Toluene-d8	10.09	98	1423574	45.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.46%	
62) 4-Bromofluorobenzene	12.40	95	389965	36.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	73.50%	
Target Compounds						
16) Acetone	3.82	43	24940	9.35	ug/l	99
52) Toluene	10.16	92	14128	0.67	ug/l	97
84) 1,2,4-Trimethylbenzene	13.04	105	5850	0.28	ug/l	99
86) p-Isopropyltoluene	13.29	119	82340	4.05	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN080618\
Data File : VN050406.D
Acq On : 7 Aug 2018 1:23
Operator : MD\SY
Sample : J4316-09
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
OW-2-4

Manual Integrations
APPROVED

MMDadoda
8/7/2018 10:46:54 AM

Quant Time: Aug 07 09:47:39 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080618W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 07 05:22:15 2018
Response via : Initial Calibration

