

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN080618\
 Data File : VN050407.D
 Acq On : 7 Aug 2018 1:48
 Operator : MD\SY
 Sample : J4355-03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-WATER

Quant Time: Aug 07 09:48:10 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080618W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 07 05:22:15 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	647508	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1043722	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	860997	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	295896	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	459887	53.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.32%	
35) Dibromofluoromethane	7.59	113	409228	48.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.72%	
50) Toluene-d8	10.09	98	1440210	45.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.48%	
62) 4-Bromofluorobenzene	12.40	95	369943	34.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	68.94%	
Target Compounds						
16) Acetone	3.83	43	19560	7.23	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN080618\
Data File : VN050407.D
Acq On : 7 Aug 2018 1:48
Operator : MD\SY
Sample : J4355-03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-WATER

Quant Time: Aug 07 09:48:10 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N080618W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 07 05:22:15 2018
Response via : Initial Calibration

