

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN080618\
 Data File : VN050410.D
 Acq On : 7 Aug 2018 3:03
 Operator : MD\SY
 Sample : J4335-03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 CINNAMINSON-8-3-18

Quant Time: Aug 07 08:01:51 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080618W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 07 05:22:15 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	649785	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1062526	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	933148	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	365201	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	455401	52.46	ug/l	0.00
Spiked Amount			Recovery	=	104.92%	
35) Dibromofluoromethane	7.59	113	411059	48.21	ug/l	0.00
Spiked Amount			Recovery	=	96.42%	
50) Toluene-d8	10.09	98	1459982	45.55	ug/l	0.00
Spiked Amount			Recovery	=	91.10%	
62) 4-Bromofluorobenzene	12.40	95	421304	38.56	ug/l	0.00
Spiked Amount			Recovery	=	77.12%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.83	43	18404	6.76	ug/l	97
18) Methyl Acetate	4.34	43	11474	1.52	ug/l	# 68
20) Methylene Chloride	4.56	84	31533	2.51	ug/l	97
30) Chloroform	7.37	83	18374	1.30	ug/l	97
43) Isopropyl Acetate	8.17	43	254674	17.55	ug/l	99
59) 2-Hexanone	10.77	43	14898	2.04	ug/l	62
95) Naphthalene	15.14	128	20527	4.70	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN080618\
Data File : VN050410.D
Acq On : 7 Aug 2018 3:03
Operator : MD\SY
Sample : J4335-03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
CINNAMINSON-8-3-18

Quant Time: Aug 07 08:01:51 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080618W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 07 05:22:15 2018
Response via : Initial Calibration

