

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN080618\
 Data File : VN050411.D
 Acq On : 7 Aug 2018 3:27
 Operator : MD\SY
 Sample : J4333-20
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SOIL-5

Quant Time: Aug 07 08:03:35 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080618W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 07 05:22:15 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	619351	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1034149	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	896283	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	369488	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	442155	53.43	ug/l	0.00
Spiked Amount						
			Recovery	=	106.86%	
35) Dibromofluoromethane	7.60	113	398960	48.08	ug/l	0.00
Spiked Amount						
			Recovery	=	96.16%	
50) Toluene-d8	10.09	98	1406053	45.07	ug/l	0.00
Spiked Amount						
			Recovery	=	90.14%	
62) 4-Bromofluorobenzene	12.40	95	412161	38.76	ug/l	0.00
Spiked Amount						
			Recovery	=	77.52%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.82	43	22930	8.94	ug/l	97
18) Methyl Acetate	4.33	43	11950	1.68	ug/l #	64
20) Methylene Chloride	4.55	84	35355	3.23	ug/l	94
25) 2-Butanone	6.86	43	4819	1.18	ug/l #	84
30) Chloroform	7.38	83	20087	1.49	ug/l	91
43) Isopropyl Acetate	8.17	43	202586	14.34	ug/l	99
59) 2-Hexanone	10.77	43	14070	1.98	ug/l #	57
95) Naphthalene	15.14	128	15543	4.43	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN080618\
Data File : VN050411.D
Acq On : 7 Aug 2018 3:27
Operator : MD\SY
Sample : J4333-20
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SOIL-5

Quant Time: Aug 07 08:03:35 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N080618W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 07 05:22:15 2018
Response via : Initial Calibration

