

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN080618\
 Data File : VN050412.D
 Acq On : 7 Aug 2018 3:52
 Operator : MD\SY
 Sample : J4333-16
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SOIL-4

Quant Time: Aug 07 08:05:26 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080618W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 07 05:22:15 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	645420	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1057383	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	925076	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	374850	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	450102	52.20	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	104.40%	
35) Dibromofluoromethane	7.60	113	409233	48.23	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	96.46%	
50) Toluene-d8	10.09	98	1457772	45.70	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	91.40%	
62) 4-Bromofluorobenzene	12.40	95	421677	38.78	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	77.56%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.83	43	23842	8.92	ug/l	99
18) Methyl Acetate	4.34	43	14619	2.00	ug/l	95
20) Methylene Chloride	4.55	84	1847032	240.76	ug/l	99
43) Isopropyl Acetate	8.17	43	202937	14.05	ug/l	98
59) 2-Hexanone	10.77	43	14147	1.95	ug/l #	59
95) Naphthalene	15.13	128	83819	7.90	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN080618\
Data File : VN050412.D
Acq On : 7 Aug 2018 3:52
Operator : MD\SY
Sample : J4333-16
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SOIL-4

Quant Time: Aug 07 08:05:26 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N080618W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 07 05:22:15 2018
Response via : Initial Calibration

