

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN080618\
 Data File : VN050420.D
 Acq On : 7 Aug 2018 7:10
 Operator : MD\SY
 Sample : J4345-03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RBR200035

Quant Time: Aug 07 08:27:15 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080618W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 07 05:22:15 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	669290	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1105404	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	934252	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	356494	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	464764	51.97	ug/l	0.00
Spiked Amount			Recovery	=	103.94%	
35) Dibromofluoromethane	7.59	113	421366	47.51	ug/l	0.00
Spiked Amount			Recovery	=	95.02%	
50) Toluene-d8	10.09	98	1495918	44.86	ug/l	0.00
Spiked Amount			Recovery	=	89.72%	
62) 4-Bromofluorobenzene	12.40	95	418273	36.80	ug/l	0.00
Spiked Amount			Recovery	=	73.60%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.83	43	25134	9.08	ug/l	98
18) Methyl Acetate	4.34	43	15768	2.09	ug/l	96
20) Methylene Chloride	4.55	84	132055	15.11	ug/l	99
30) Chloroform	7.37	83	15136	1.04	ug/l	91
43) Isopropyl Acetate	8.17	43	233453	15.46	ug/l	98
59) 2-Hexanone	10.77	43	11513	1.52	ug/l #	58
95) Naphthalene	15.14	128	71165	7.44	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN080618\
Data File : VN050420.D
Acq On : 7 Aug 2018 7:10
Operator : MD\SY
Sample : J4345-03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 49 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RBR200035

Quant Time: Aug 07 08:27:15 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N080618W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 07 05:22:15 2018
Response via : Initial Calibration

