

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN080619\
 Data File : VN057096.D
 Acq On : 6 Aug 2019 12:24
 Operator : JC/SP
 Sample : K4131-32
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 1S-E1-(0-1)

Quant Time: Aug 07 04:01:25 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072719W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jul 27 01:33:29 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	590696	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	855213	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	704540	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	241622	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	291756	48.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.50%	
35) Dibromofluoromethane	7.58	113	253405	46.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.58%	
50) Toluene-d8	10.08	98	976276	46.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.52%	
62) 4-Bromofluorobenzene	12.40	95	281640	41.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	82.88%	
Target Compounds						
					Qvalue	
18) Methyl Acetate	4.32	43	15890	2.274	ug/l	92
20) Methylene Chloride	4.52	84	7152	1.133	ug/l	90
43) Isopropyl Acetate	8.17	43	35569	3.571	ug/l #	72
95) Naphthalene	15.12	128	4119	3.960	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN080619\
Data File : VN057096.D
Acq On : 6 Aug 2019 12:24
Operator : JC/SP
Sample : K4131-32
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
1S-E1-(0-1)

Quant Time: Aug 07 04:01:25 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N072719W.M
Quant Title : SW846 8260
QLast Update : Sat Jul 27 01:33:29 2019
Response via : Initial Calibration

