

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN080619\
 Data File : VN057098.D
 Acq On : 6 Aug 2019 13:11
 Operator : JC/SP
 Sample : K4131-22
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 1S-D1-(0-1)

Quant Time: Aug 07 04:09:06 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072719W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jul 27 01:33:29 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	521269	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	749551	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	630906	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	210551	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	260545	48.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.66%	
35) Dibromofluoromethane	7.58	113	224254	46.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.48%	
50) Toluene-d8	10.08	98	872334	47.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.34%	
62) 4-Bromofluorobenzene	12.40	95	252663	42.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	84.82%	
Target Compounds						
16) Acetone	3.82	43	10515	5.439	ug/l	92
18) Methyl Acetate	4.32	43	17954	2.911	ug/l	99
20) Methylene Chloride	4.52	84	7060	1.268	ug/l	88
43) Isopropyl Acetate	8.17	43	56282	6.446	ug/l #	65

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN080619\
Data File : VN057098.D
Acq On : 6 Aug 2019 13:11
Operator : JC/SP
Sample : K4131-22
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
1S-D1-(0-1)

Quant Time: Aug 07 04:09:06 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N072719W.M
Quant Title : SW846 8260
QLast Update : Sat Jul 27 01:33:29 2019
Response via : Initial Calibration

