

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN080619\
 Data File : VN057100.D
 Acq On : 6 Aug 2019 13:58
 Operator : JC/SP
 Sample : K4152-32
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 2S-E1-(0-1)

Quant Time: Aug 07 04:14:53 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072719W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jul 27 01:33:29 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	502696	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	725760	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	608910	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	203618	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	248724	48.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.68%	
35) Dibromofluoromethane	7.58	113	216094	46.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.02%	
50) Toluene-d8	10.08	98	843112	47.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.18%	
62) 4-Bromofluorobenzene	12.40	95	240314	41.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	83.32%	
Target Compounds						
8) Diethyl Ether	3.40	74	3707	1.326	ug/l	80
16) Acetone	3.82	43	15346	8.232	ug/l #	85
18) Methyl Acetate	4.32	43	21101	3.548	ug/l	100
20) Methylene Chloride	4.53	84	15175	2.826	ug/l	87
43) Isopropyl Acetate	8.17	43	49967	5.911	ug/l #	68

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN080619\
Data File : VN057100.D
Acq On : 6 Aug 2019 13:58
Operator : JC/SP
Sample : K4152-32
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
2S-E1-(0-1)

Quant Time: Aug 07 04:14:53 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N072719W.M
Quant Title : SW846 8260
QLast Update : Sat Jul 27 01:33:29 2019
Response via : Initial Calibration

