

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN080619\
 Data File : VN057101.D
 Acq On : 6 Aug 2019 14:21
 Operator : JC/SP
 Sample : K4152-27
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 2S-H1-(30)

Manual Integrations
 APPROVED

MMdadoda
 8/9/2019 12:51:35 AM

Quant Time: Aug 07 04:16:44 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072719W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jul 27 01:33:29 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	494265	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	712065	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	607136	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	192465	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	247939	49.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.00%	
35) Dibromofluoromethane	7.58	113	216037	47.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.78%	
50) Toluene-d8	10.08	98	830975	47.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.60%	
62) 4-Bromofluorobenzene	12.40	95	239095	42.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	84.50%	
Target Compounds						
8) Diethyl Ether	3.40	74	3474m	1.264	ug/l	Qvalue
16) Acetone	3.81	43	15064	8.218	ug/l	95
18) Methyl Acetate	4.31	43	18480	3.160	ug/l	97
20) Methylene Chloride	4.52	84	16278	3.083	ug/l	91
43) Isopropyl Acetate	8.16	43	47074	5.675	ug/l #	68

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN080619\
Data File : VN057101.D
Acq On : 6 Aug 2019 14:21
Operator : JC/SP
Sample : K4152-27
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
2S-H1-(30)

Manual Integrations
APPROVED

MMdadoda
8/9/2019 12:51:35 AM

Quant Time: Aug 07 04:16:44 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072719W.M
Quant Title : SW846 8260
QLast Update : Sat Jul 27 01:33:29 2019
Response via : Initial Calibration

