


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN080619\
Data File : VN057089.D
Acq On : 6 Aug 2019 8:47
Operator : JC/SP
Sample : BFB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072719W.M
Title : SW846 8260
Last Update : Sat Jul 27 01:33:29 2019



AutoFind: Scans 3303, 3304, 3305; Background Corrected with Scan 3289

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	15.9	10335	PASS
75	95	30	60	46.0	30013	PASS
95	95	100	100	100.0	65178	PASS
96	95	5	9	6.6	4294	PASS
173	174	0.00	2	1.2	711	PASS
174	95	50	100	94.8	61821	PASS
175	174	5	9	7.7	4765	PASS
176	174	95	101	95.1	58821	PASS
177	176	5	9	7.0	4115	PASS