

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN080620\
 Data File : VN062766.D
 Acq On : 6 Aug 2020 15:05
 Operator : JC/MD
 Sample : PB130737TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB130737TB

Manual Integrations
 APPROVED

MMDadoda
 8/7/2020 12:50:46 PM

Quant Time: Aug 07 03:51:43 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080520W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 05 13:26:42 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	132740	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	232090	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	238206	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	83841	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	106336	48.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.88%	
35) Dibromofluoromethane	7.55	113	77757	48.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.96%	
50) Toluene-d8	10.06	98	305223	49.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.48%	
62) 4-Bromofluorobenzene	12.38	95	112303	47.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.90%	
Target Compounds						
16) Acetone	3.78	43	22248	28.049	ug/l	99
18) Methyl Acetate	4.30	43	5278m	1.723	ug/l	
20) Methylene Chloride	4.50	84	8440	5.132	ug/l	96
43) Isopropyl Acetate	8.13	43	54070	13.732	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN080620\
Data File : VN062766.D
Acq On : 6 Aug 2020 15:05
Operator : JC/MD
Sample : PB130737TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB130737TB

Manual Integrations
APPROVED

MMDadoda
8/7/2020 12:50:46 PM

Quant Time: Aug 07 03:51:43 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N080520W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 05 13:26:42 2020
Response via : Initial Calibration

