

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN080620\
 Data File : VN062772.D
 Acq On : 6 Aug 2020 17:31
 Operator : JC/MD
 Sample : L3553-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 JB-1A

Quant Time: Aug 07 04:38:34 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080520W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 05 13:26:42 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	132341	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	242048	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	247652	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	89275	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	112336	51.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.66%	
35) Dibromofluoromethane	7.55	113	78560	47.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.90%	
50) Toluene-d8	10.06	98	315545	48.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.62%	
62) 4-Bromofluorobenzene	12.38	95	114253	46.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.54%	
Target Compounds						
16) Acetone	3.78	43	18763	23.727	ug/l	98
18) Methyl Acetate	4.28	43	5634	1.926	ug/l #	79
20) Methylene Chloride	4.50	84	9077	5.597	ug/l	96
43) Isopropyl Acetate	8.13	43	44453	10.825	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN080620\
Data File : VN062772.D
Acq On : 6 Aug 2020 17:31
Operator : JC/MD
Sample : L3553-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
JB-1A

Quant Time: Aug 07 04:38:34 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N080520W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 05 13:26:42 2020
Response via : Initial Calibration

