

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN080620\
 Data File : VN062775.D
 Acq On : 6 Aug 2020 18:44
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 07 04:53:32 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080520W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 05 13:26:42 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	117	0.00
2 T	Dichlorodifluoromethane	0.716	0.729	-1.8	112	0.00
3 P	Chloromethane	0.676	0.770	-13.9	137	0.00
4 C	Vinyl Chloride	0.634	0.721	-13.7#	130	0.00
5 T	Bromomethane	0.419	0.377	10.0	113	0.00
6 T	Chloroethane	0.417	0.413	1.0	117	0.00
7 T	Trichlorofluoromethane	1.130	1.052	6.9	112	0.00
8 T	Diethyl Ether	0.323	0.301	6.8	110	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.502	0.454	9.6	115	0.00
10 T	Methyl Iodide	0.548	0.558	-1.8	113	0.00
11 T	Tert butyl alcohol	0.136	0.114	16.2	96	0.00
12 CM	1,1-Dichloroethene	0.476	0.414	13.0#	111	0.00
13 T	Acrolein	0.082	0.069	15.9	101	0.00
14 T	Allyl chloride	0.875	0.811	7.3	116	0.00
15 T	Acrylonitrile	0.314	0.300	4.5	108	0.00
16 T	Acetone	0.299	0.251	16.1	103	0.00
17 T	Carbon Disulfide	1.404	1.139	18.9	110	0.00
18 T	Methyl Acetate	0.794	0.692	12.8	110	0.00
19 T	Methyl tert-butyl Ether	1.937	1.835	5.3	108	0.00
20 T	Methylene Chloride	0.609	0.520	14.6	110	0.00
21 T	trans-1,2-Dichloroethene	0.500	0.448	10.4	110	0.00
22 T	Diisopropyl ether	1.864	1.881	-0.9	115	0.00
23 T	Vinyl Acetate	1.417	1.316	7.1	111	0.00
24 P	1,1-Dichloroethane	1.092	0.989	9.4	110	0.00
25 T	2-Butanone	0.423	0.393	7.1	105	0.00
26 T	2,2-Dichloropropane	1.007	0.745	26.0#	92	0.00
27 T	cis-1,2-Dichloroethene	0.600	0.556	7.3	113	0.00
28 T	Bromochloromethane	0.527	0.508	3.6	118	0.00
29 T	Tetrahydrofuran	0.275	0.260	5.5	105	0.00
30 C	Chloroform	1.161	1.025	11.7#	107	0.00
31 T	Cyclohexane	1.002	0.949	5.3	115	0.00
32 T	1,1,1-Trichloroethane	1.038	0.914	11.9	104	0.00
33 S	1,2-Dichloroethane-d4	0.827	0.777	6.0	115	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	118	0.00
35 S	Dibromofluoromethane	0.342	0.331	3.2	117	0.00
36 T	1,1-Dichloropropene	0.487	0.442	9.2	108	0.00
37 T	Ethyl Acetate	0.509	0.458	10.0	106	0.00
38 T	Carbon Tetrachloride	0.576	0.495	14.1	107	0.00
39 T	Methylcyclohexane	0.586	0.566	3.4	115	0.00
40 TM	Benzene	1.453	1.303	10.3	110	0.00
41 T	Methacrylonitrile	0.229	0.239	-4.4	111	0.00
42 TM	1,2-Dichloroethane	0.618	0.546	11.7	106	0.00
43 T	Isopropyl Acetate	0.848	0.772	9.0	108	0.00
44 TM	Trichloroethene	0.354	0.320	9.6	112	0.00
45 C	1,2-Dichloropropane	0.393	0.366	6.9#	110	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN080620\
 Data File : VN062775.D
 Acq On : 6 Aug 2020 18:44
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 07 04:53:32 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080520W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 05 13:26:42 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.258	0.242	6.2	111	0.00
47 T	Bromodichloromethane	0.579	0.535	7.6	107	0.00
48 T	Methyl methacrylate	0.417	0.386	7.4	107	0.00
49 T	1,4-Dioxane	0.008	0.007	12.5	98	0.00
50 S	Toluene-d8	1.335	1.286	3.7	116	0.00
51 T	4-Methyl-2-Pentanone	0.549	0.518	5.6	107	0.00
52 CM	Toluene	0.897	0.830	7.5#	109	0.00
53 T	t-1,3-Dichloropropene	0.636	0.563	11.5	102	0.00
54 T	cis-1,3-Dichloropropene	0.639	0.588	8.0	108	0.00
55 T	1,1,2-Trichloroethane	0.383	0.343	10.4	110	0.00
56 T	Ethyl methacrylate	0.577	0.564	2.3	107	0.00
57 T	1,3-Dichloropropane	0.651	0.608	6.6	109	0.00
58 T	2-Chloroethyl Vinyl ether	0.163	0.150	8.0	107	0.00
59 T	2-Hexanone	0.409	0.386	5.6	105	0.00
60 T	Dibromochloromethane	0.437	0.404	7.6	108	0.00
61 T	1,2-Dibromoethane	0.378	0.348	7.9	109	0.00
62 S	4-Bromofluorobenzene	0.505	0.489	3.2	115	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	117	0.00
64 T	Tetrachloroethene	0.323	0.296	8.4	109	0.00
65 PM	Chlorobenzene	1.045	0.977	6.5	110	0.00
66 T	1,1,1,2-Tetrachloroethane	0.420	0.384	8.6	106	0.00
67 C	Ethyl Benzene	1.870	1.800	3.7#	109	0.00
68 T	m/p-Xylenes	0.679	0.658	3.1	108	0.00
69 T	o-Xylene	0.636	0.630	0.9	107	0.00
70 T	Styrene	1.118	1.123	-0.4	108	0.00
71 P	Bromoform	0.343	0.311	9.3	103	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	112	0.00
73 T	Isopropylbenzene	3.765	3.631	3.6	107	0.00
74 T	N-amyl acetate	1.526	1.566	-2.6	107	0.00
75 P	1,1,2,2-Tetrachloroethane	1.325	1.198	9.6	107	0.00
76 T	1,2,3-Trichloropropane	1.348	1.175	12.8	102	0.00
77 T	Bromobenzene	0.935	0.871	6.8	109	0.00
78 T	n-propylbenzene	4.420	4.284	3.1	107	0.00
79 T	2-Chlorotoluene	2.771	2.557	7.7	104	0.00
80 T	1,3,5-Trimethylbenzene	3.266	3.212	1.7	106	0.00
81 T	trans-1,4-Dichloro-2-butene	0.442	0.387	12.4	95	0.00
82 T	4-Chlorotoluene	2.856	2.710	5.1	107	0.00
83 T	tert-Butylbenzene	2.764	2.685	2.9	107	0.00
84 T	1,2,4-Trimethylbenzene	3.125	3.176	-1.6	108	0.00
85 T	sec-Butylbenzene	3.629	3.540	2.5	107	0.00
86 T	p-Isopropyltoluene	3.222	3.194	0.9	106	0.00
87 T	1,3-Dichlorobenzene	1.671	1.556	6.9	104	0.00
88 T	1,4-Dichlorobenzene	1.676	1.547	7.7	106	0.00
89 T	n-Butylbenzene	2.725	2.663	2.3	107	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN080620\
Data File : VN062775.D
Acq On : 6 Aug 2020 18:44
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Aug 07 04:53:32 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080520W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 05 13:26:42 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.694	0.609	12.2	103	0.00
91 T	1,2-Dichlorobenzene	1.567	1.525	2.7	108	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.285	0.257	9.8	98	0.00
93 T	1,2,4-Trichlorobenzene	0.734	0.721	1.8	106	0.00
94 T	Hexachlorobutadiene	0.545	0.467	14.3	102	0.00
95 T	Naphthalene	1.932	1.927	0.3	105	0.00
96 T	1,2,3-Trichlorobenzene	0.714	0.723	-1.3	105	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6