

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN080718\
 Data File : VN050434.D
 Acq On : 7 Aug 2018 16:15
 Operator : MD\SY
 Sample : J4297-08
 Misc : 5.06g/5mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-8

Manual Integrations
 APPROVED

MMDadoda
 8/8/2018 10:47:43 AM

Quant Time: Aug 08 06:55:02 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080618W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 07 05:22:15 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	819116	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1235197	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1084332	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	648412	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	484605	44.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.56%	
35) Dibromofluoromethane	7.59	113	438519	44.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.50%	
50) Toluene-d8	10.09	98	1757842	47.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.34%	
62) 4-Bromofluorobenzene	12.40	95	642608	50.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.20%	
Target Compounds						
10) Methyl Iodide	3.93	142	10374	2.51	ug/l	Qvalue 91
16) Acetone	3.84	43	4117m	0.90	ug/l	
18) Methyl Acetate	4.34	43	17434	1.87	ug/l	98
39) Methylcyclohexane	9.08	83	29497	1.99	ug/l	95
85) sec-Butylbenzene	13.17	105	83664	1.82	ug/l #	73

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN080718\
Data File : VN050434.D
Acq On : 7 Aug 2018 16:15
Operator : MD\SY
Sample : J4297-08
Misc : 5.06g/5mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-8

Manual Integrations
APPROVED

MMDadoda
8/8/2018 10:47:43 AM

Quant Time: Aug 08 06:55:02 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N080618W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 07 05:22:15 2018
Response via : Initial Calibration

