

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN080718\
 Data File : VN050441.D
 Acq On : 7 Aug 2018 19:09
 Operator : MD\SY
 Sample : J4369-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-20

Manual Integrations
 APPROVED

MMDadoda
 8/8/2018 10:48:17 AM

Quant Time: Aug 13 08:16:47 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080618W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 07 05:22:15 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	803537	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1217914	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1059985	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	468537	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	510533	47.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.10%	
35) Dibromofluoromethane	7.59	113	460355	47.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.22%	
50) Toluene-d8	10.09	98	1679094	45.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.40%	
62) 4-Bromofluorobenzene	12.40	95	535064	42.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.46%	
Target Compounds						
11) Tert butyl alcohol	4.80	59	24176	28.03	ug/l	99
16) Acetone	3.83	43	19116	5.62	ug/l	100
19) Methyl tert-butyl Ether	5.05	73	270911	10.99	ug/l	99
65) Chlorobenzene	11.44	112	23368	0.88	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN080718\
Data File : VN050441.D
Acq On : 7 Aug 2018 19:09
Operator : MD\SY
Sample : J4369-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-20

Manual Integrations
APPROVED

MMDadoda
8/8/2018 10:48:17 AM

Quant Time: Aug 13 08:16:47 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080618W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 07 05:22:15 2018
Response via : Initial Calibration

