

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN080718\
 Data File : VN050445.D
 Acq On : 7 Aug 2018 20:48
 Operator : MD\SY
 Sample : J4321-01
 Misc : 6.09g/5.0mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 EP1-NW-IMPACTED-0818

Manual Integrations
 APPROVED

MMDadoda
 8/8/2018 10:48:34 AM

Quant Time: Aug 08 10:59:37 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080618W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 07 05:22:15 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	809837	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1211395	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1066320	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	543030	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	489958	45.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.56%	
35) Dibromofluoromethane	7.59	113	444283	45.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.42%	
50) Toluene-d8	10.09	98	1737560	47.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.08%	
62) 4-Bromofluorobenzene	12.40	95	581175	46.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.32%	
Target Compounds						
39) Methylcyclohexane	9.08	83	89522	6.16	ug/l	99
73) Isopropylbenzene	12.25	105	416354	10.21	ug/l	100
78) n-propylbenzene	12.59	91	936284	20.29	ug/l	100
85) sec-Butylbenzene	13.17	105	458266	11.91	ug/l #	78
89) n-Butylbenzene	13.62	91	661696	23.57	ug/l #	77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN080718\
Data File : VN050445.D
Acq On : 7 Aug 2018 20:48
Operator : MD\SY
Sample : J4321-01
Misc : 6.09g/5.0mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
EP1-NW-IMPACTED-0818

Quant Time: Aug 08 10:59:37 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080618W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 07 05:22:15 2018
Response via : Initial Calibration

Manual Integrations
APPROVED

MMDadoda
8/8/2018 10:48:34 AM

