

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN080718\
 Data File : VN050446.D
 Acq On : 7 Aug 2018 21:13
 Operator : MD\SY
 Sample : PB111815TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB111815TB

Quant Time: Aug 08 07:37:38 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080618W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 07 05:22:15 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	874977	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1340060	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1181390	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	613538	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	530496	45.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.76%	
35) Dibromofluoromethane	7.59	113	488002	45.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.76%	
50) Toluene-d8	10.09	98	1890281	46.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.52%	
62) 4-Bromofluorobenzene	12.40	95	605525	43.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.90%	
Target Compounds						
16) Acetone	3.83	43	20414	5.50	ug/l	94
18) Methyl Acetate	4.33	43	13736	1.33	ug/l	96
43) Isopropyl Acetate	8.17	43	339140	18.53	ug/l	98
59) 2-Hexanone	10.77	43	28576	3.11	ug/l #	59

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN080718\
Data File : VN050446.D
Acq On : 7 Aug 2018 21:13
Operator : MD\SY
Sample : PB111815TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampled :
PB111815TB

Quant Time: Aug 08 07:37:38 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N080618W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 07 05:22:15 2018
Response via : Initial Calibration

