

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN080720\
 Data File : VN062782.D
 Acq On : 7 Aug 2020 16:28
 Operator : JC/MD
 Sample : PB130807TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB130807TB

Quant Time: Aug 08 02:39:33 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080520W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 05 13:26:42 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	138779	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	239378	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	239861	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	86972	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	112881	49.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.38%	
35) Dibromofluoromethane	7.55	113	81760	49.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.86%	
50) Toluene-d8	10.06	98	310248	48.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.06%	
62) 4-Bromofluorobenzene	12.38	95	110652	45.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.60%	
Target Compounds						
20) Methylene Chloride	4.50	84	6707	3.716	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN080720\
Data File : VN062782.D
Acq On : 7 Aug 2020 16:28
Operator : JC/MD
Sample : PB130807TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB130807TB

Quant Time: Aug 08 02:39:33 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N080520W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 05 13:26:42 2020
Response via : Initial Calibration

