

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN080720\
 Data File : VN062783.D
 Acq On : 7 Aug 2020 16:52
 Operator : JC/MD
 Sample : PB130790TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB130790TB

Quant Time: Aug 08 02:41:12 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080520W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 05 13:26:42 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	131115	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	229719	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	240045	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	83150	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	108590	50.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.16%	
35) Dibromofluoromethane	7.55	113	80559	51.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.54%	
50) Toluene-d8	10.06	98	308546	50.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.58%	
62) 4-Bromofluorobenzene	12.38	95	108004	46.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.18%	
Target Compounds						
16) Acetone	3.78	43	10826	13.818	ug/l	94
18) Methyl Acetate	4.29	43	4082	1.098	ug/l #	81
20) Methylene Chloride	4.50	84	12801	8.294	ug/l	94
43) Isopropyl Acetate	8.13	43	47980	12.311	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN080720\
Data File : VN062783.D
Acq On : 7 Aug 2020 16:52
Operator : JC/MD
Sample : PB130790TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB130790TB

Quant Time: Aug 08 02:41:12 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N080520W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 05 13:26:42 2020
Response via : Initial Calibration

