

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN080720\
 Data File : VN062793.D
 Acq On : 7 Aug 2020 20:55
 Operator : JC/MD
 Sample : L3537-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 JTY1-WC-L-01

Quant Time: Aug 10 08:02:17 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080520W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 05 13:26:42 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	131555	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	246102	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	237134	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	90416	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	111213	51.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.24%	
35) Dibromofluoromethane	7.55	113	81185	48.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.44%	
50) Toluene-d8	10.06	98	310252	47.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.42%	
62) 4-Bromofluorobenzene	12.38	95	111526	44.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.80%	
Target Compounds						
					Qvalue	
3) Chloromethane	2.03	50	1139	0.641	ug/l #	76
16) Acetone	3.78	43	41796	53.169	ug/l	99
20) Methylene Chloride	4.50	84	7296	4.378	ug/l	98
25) 2-Butanone	6.80	43	7482	6.720	ug/l #	86
40) Benzene	8.00	78	268831	37.595	ug/l	100
52) Toluene	10.13	92	22994	5.208	ug/l	100
67) Ethyl Benzene	11.49	91	4498	0.507	ug/l #	88
68) m/p-Xylenes	11.60	106	2115	0.657	ug/l	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN080720\
Data File : VN062793.D
Acq On : 7 Aug 2020 20:55
Operator : JC/MD
Sample : L3537-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
JTY1-WC-L-01

Quant Time: Aug 10 08:02:17 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N080520W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 05 13:26:42 2020
Response via : Initial Calibration

