

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN080818\
 Data File : VN050460.D
 Acq On : 8 Aug 2018 15:39
 Operator : MD\SY
 Sample : PB111846TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB111846TB

Quant Time: Aug 09 01:17:37 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080618W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 07 05:22:15 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	887997	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1355058	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1177063	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	460466	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	549815	46.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.68%	
35) Dibromofluoromethane	7.59	113	507080	46.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.28%	
50) Toluene-d8	10.09	98	1872917	45.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.62%	
62) 4-Bromofluorobenzene	12.40	95	550591	39.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	79.04%	
Target Compounds						
16) Acetone	3.82	43	24067	6.45	ug/l	99
18) Methyl Acetate	4.33	43	11847	1.10	ug/l #	87
20) Methylene Chloride	4.55	84	32053	1.46	ug/l	95
43) Isopropyl Acetate	8.17	43	314927	17.01	ug/l	97
95) Naphthalene	15.14	128	14079	4.21	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN080818\
Data File : VN050460.D
Acq On : 8 Aug 2018 15:39
Operator : MD\SY
Sample : PB111846TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB111846TB

Quant Time: Aug 09 01:17:37 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N080618W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 07 05:22:15 2018
Response via : Initial Calibration

